HP Archive

:

. . .

. *

· . .

This vintage Hewlett Packard document was preserved and distributed by

www. hparchive.com

Please visit us on the web !

On-line curator: Tony Gerbic

 a^{3}

OPERATING AND SERVICING MANUAL



12

Cherry

-15

-

10

the state of the second

大学

Constant and

MODEL 739A FREQUENCY RESPONSE TEST SET SERIALS PREFIXED: 944 -010-00483



Copyright HEWLETT-PACKARD COMPANY 1959 275 PAGE MILL ROAD, PALO ALTO, CALIFORNIA, U.S.A.

SECTION I Description

The *m* Model 739A Frequency Response Test Set simplifies frequency response determination by providing a convenient constant amplitude reference voltage of variable frequency. Frequency response of vacuum tube voltmeters, oscilloscope, video amplifiers and filters from 5 cps to 10 mc* can be quickly and easily checked.

The Frequency Response Test Set contains a wide range oscillator, special voltmeter and step attenuator as shown in Figure 1. The oscillator generates constant amplitude signals between 300 kc and 10 mc. The monitoring circuit is flat from 5 cps to 10 mc so that an external oscillator such as \bigcirc Model 200S can be used to reduce the lower frequency limit to as low as 5 cps. Oscillator output is applied to the monitoring voltmeter, which samples the input to a step attenuator. An arbitrary reference level is maintained on the monitoring meter as the frequency setting is varied. The attenuator reduces the signal to an appropriate output level which appears across 50 ohms at the end of the special output cable supplied.





SPECIFICATIONS

FREQUENCY RANGE:

INTERNAL OSCILLATOR:

REQUIREMENTS FOR EXTERNAL OSCILLATOR:

FREQUENCY RESPONSE OF MONITORING CIRCUIT:

OUTPUT:

DIMENSIONS:

ACCESSORIES FURNISHED:

5 cycles* to 10 mc.

300 kc to 10 mc in 3 ranges.

Frequency range 5 cps to 300 kc requires 3 volts into 50 ohms, distortion less than 1%.

Flat within $\pm 0.5\%$ from 10 cps to 5 mc; within $\pm 0.5\%$, -1.5%, 5 cps to 10 mc. Monitor circuit is average responding.

At least 3 volts across 50 ohm cable termination. Adjustable in 10 db steps by a 0 to 70 db attenuator. Fine adjustment provided.

Rack Mount: 7 in. high, 19 in. wide, 9 in. deep.

739A-16A Output Cable, 50 ohm termination. BNC to shielded dual banana plug, (extra cables available on special order).

* Extended frequency range when used with @ Model 200S Oscillator.

SECTION II OPERATION

Calibrate the device under test with an accurate voltage standard, such as the Model 738A Voltmeter Calibrator. This voltage standard should have an output frequency between 100 and 1000 cps.

1) Connect the device under test to the Frequency Response Test Set OUTPUT connector, using the special terminated output cable provided.

2) Connect a suitable oscillator, such as the PModel 200S Oscillator, to the EXTERNAL INPUT connector on the Frequency Response Test Set.

3) Set the external oscillator to the frequency used for calibration.

4) Set the RANGE SELECTOR to EXTERNAL. Set a reference on the device under test using the external oscillator AMPLITUDE control and the OUTPUT ATTENUATOR of the Test Set.

NOTE

Always use the OUTPUT ATTENUATOR range that corresponds to the desired input to the device under test. This will assure a Test Set meter indication near SET LEVEL.

5) Set the meter in the Test Set to SET LEVEL with the METER SET control.

6) Shift to any other frequency within the range of the Test Set or external oscillator. Return the meter to SET LEVEL using the oscillator AMPLITUDE control. Note any deviation from the reference set in step 4 on the device under test.

7) Repeat steps 3 through 6 if you change the range switch on the device under test. Repeat step 6 for all frequencies you desire to check.

SECTION III MAINTENANCE

The internal oscillator is a single pentode (V2) located inside the shield on top of the chassis. The oscillator plate and screen voltages are controlled by a series regulator triode (V1). This series regulator is controlled by a two stage differential amplifier (V3 and V4) that senses the amplitude of the voltage applied to the OUTPUT ATTENUATOR. The combined action of the series regulator, differential amplifier and output amplitude sensing circuits keeps the output constant over the range of the internal oscillator.

The power supply circuit is conventional, using

glow discharge tubes to regulate the differential amplifier voltages.

Tube element voltages are not listed on the schematic diagram since they vary radically with frequency and output AMPLITUDE control setting.

The only adjustments in the oscillator circuit are for initial dial calibration and should not need further attention. Another adjustment (C16) is provided in the meter circuit. This meter frequency response adjustment should not be touched unless you have accurate means of checking the accuracy of the reference meter between 7 and 10 mc.





SECTION IV TABLE OF REPLACEABLE PARTS

44

NOTE -

Readily available standard-components have been used in this instrument, whenever possible. However, special components may be obtained from your local Hewlett-Packard representative or from the factory.

When ordering parts always include:

- 1. 🖗 Stock Number.
- 2. Complete description of part including circuit reference.
- 3. Model number and serial number of instrument.
- 4. If part is not listed give complete description, function, and location of part.

If there are any corrections for the Table of Replaceable Parts they will be listed on an Instruction Manual Change sheet at the front of this manual.

9

CIRCUIT REF.	DESCRIPTION, MFR. * & MFR. DESIGNATIO	ON	STOCK NO.	#			
C1	Capacitor: variable, air dielectric, 2 section, 9.5 - 443.2 $\mu\mu f$	w*	12-39	1			
C2	Capacitor: variable, ceramic, 7 - 45 μμf, 500 vdcw	L*	13-1	1		x i	
C3, 4	Capacitor: variable, ceramic, 1 section 1.5 - 7 $\mu\mu$ f, 500 vdcw	L*	13-7	2	× 8	1	
C5	Capacitor: fixed, ceramic, . 01 μ f -0% + 100%, 1000 vdcw	C*	15-43	10			
C6	Capacitor: fixed, ceramic, 1000 $\mu\mu$ f ±25%, 500 vdcw	K*	15-21	1			
C7	Same as C5						
C8	Capacitor: fixed, mica, 100 $\mu\mu$ f ±10%, 300 vdcw	v*	14-76	1		10	
C9 thru C11	Same as C5				- 7 - 8 1		
C12, 13, 14, 15	Capacitor: fixed, mica, 820 $\mu\mu$ f, \pm 10%, 500 vdcw	v*	14-28	4			
C16	Capacitor: variable, plastic, 0.7 - $\mu\mu$ f, 350 vdcw	L*	13-26	1			
C17	Capacitor: fixed, titanium dioxide, l μμf ±10%, 500 vdcw Electrical value adjusted at the factory D	D*	15-102	1	2.		
C18, 19	Same as C5						
C20, 21	Capacitor: electrolytic, 2 sections, 20 μ f/sect., 450 vdcw C	C*	18-22HP				
C22 , 23, 24	Same as C5						
CRI, 2	Diode, germanium: HD-2135 B	U*	212-G11A	2			
CR3, 4	Diode, silicon H	P *	G-29B-5	2			
DS1	Lamp, miniature: 2 pin base, #12	0*	211-78	1			

TABLE OF REPLACEABLE PARTS

* See "List of Manufacturers Code Letters For Replaceable Parts Table". # Total quantity used in the instrument.

5

CIRCUIT REF.	DESCRIPTION, MFR. * & MFR. DESIG	GNATION	STOCK NO.	#			
Fl	Fuse: slow blow, for 115 V operation Fuse: 1/2 amp, slow blow, for 230 volt operation	E*	211-18 211-20	1 1			
Jl	Connector: type BNC, UG291/U	LL*	125-UG 291/U	1	1		
J2	Part of Output Attenuator (not separately replaceable)	LL*	1	-			
Ll, 2	Inductor, RF; 10 μ h ±10%	CG*	48-54	2			
M 1	Meter	BF HP*	112- 26 //4	1			
Pl	Cable, power	HP*	812-56	1		×	
Rl	Resistor: fixed, composition, 33 ohms ±10%, 1 W	В*	24-33	1			
R2	Resistor: fixed, composition, 1000 ohms ±10%, 1 W	в*	24-1K	1			
R3	Resistor: fixed, composition, 1000 ohms ±10%, 1/2 W	B*	23-1K	1			
R4	Resistor: fixed, composition, 82,000 ohms ±10%, 1 W	B*	24- 82K 68K	1			
R5	<pre>&% ₭ Resistor: fixed, composition, 1 megohm ±10%, 1/2 W</pre>	в*	23-1M	1			
R6	Resistor: fixed, deposited carbon, 667,000 ohms ±1%, 1/2 W	NN*	33-667K	2			
R7	Resistor: fixed, composition, 470,000 ohms ±10%, 1/2 W	B*	23-470K	1			
R8	Resistor: fixed, deposited carbon, 4 megohms ±1%, 1/2 W	NN*	33-4M	2			
R9	Resistor: fixed, deposited carbon, 5.55 megohms ±1%, 1/2 W	NN*	33-5. 55M	2			
R10	Resistor: fixed, composition, 220,000 ohms ±10%, 1/2 W	В*	23-220K	1			

TABLE OF REPLACEABLE PARTS

* See "List of Manufacturers Code Letters For Replaceable Parts Table". # Total quantity used in the instrument.

CIRCUIT REF.	DESCRIPTION, MFR. * & MFR. DESIGNATION	STOCK NO.	#	 	
R11	Same as R6				
R12	Same as R8				
R13	Same as R9			101	
R14	Resistor: fixed, deposited carbon, 600,000 ohms ±1%, 1/2 W NN*	33-600K	1	10	
R15	Resistor: fixed, deposited carbon, 2 megohms ±1%, 1/2 W NN*	33-2M	1		. 12
R16	Resistor: variable, composition, 100,000 ohms, linear taper BO*	210-82	1		
R17	Resistor: fixed, composition, 47,000 ohms ±10%, 1/2 W B*	23-47K	1		22
R18, 19	Resistor: fixed, composition, l megohm ±10%, 1/2 W B*	23-1M	1		
R20, 21	Resistor: fixed, deposited carbon, 5030 ohms ±10%, 1/2 W NN*	33-5. 03KR	1		
R22	Resistor: variable, composition, linear taper 2000 ohms ±20%, 1/4 W I*	210-4-25	1		
R23	Resistor: fixed, deposited carbon, 300 ohms ±1%, 1/2 W NN*	33-300	1		
R24, 25	Resistor: fixed, deposited carbon, 201 ohms ±1%, 1/2 W NN*	33-201	2		
R26, 27	Resistor: fixed, composition, 12,000 ohms ±10%, 2 W B*	25-12K	2		
S1	Switch HP*	Special			
S2	Switch, toggle: SPST, 250 V, 3 amp CR*	310-11			
SRI	Rectifier, selenium: 162V rms max, 20 ma dc max CN*	212-114			
Tl thru					
Т3	Transformer, RF HP*	Special	1		
Τ4	Transformer, power HP*	910-80	, 1		
					1 V (1

TABLE OF REPLACEABLE PARTS

* See "List of Manufacturers Code Letters For Replaceable Parts Table". # Total quantity used in the instrument.

steam selected at factory. Average value shown. ** optimum value

CIRCUIT

REF.

V1

V2

V5

VR1

VR4

VR2, 3

V3, 4

1

z

TABLE OF REPLAC	CEABLE	E PARTS				
DESCRIPTION, MFR. * & MFR. DESIG	NATION	STOCK NO.	#			
Tube, electron: 6C4	0*	212-6C4	1			
Tube, electron: 6AQ5	Y*	212-6AQ5	1			,
Tube, electron: 12AX7	Y*	212-12AX7	2			
Tube, electron: 6X4	Ý*	212-6X4	1	1		
Voltage regulator: OA2	BN*	212-0A2	2			
Voltage regulator: OB2	Y*	212-OB2	1			
Same as VR1						
MISCELLANEOUS Fuseholder: extractor post type	Т*	1400-002 140-16-	4 ≉ 1			
Knob: METER SET	HP*	G-74C	2			
Knob: AMPLITUDE	HP*	G-74H	1			
Knob: RANGE SELECTOR, OUTPUT ATTENUATOR	HP*	G-74N	2			
Output Cable: terminated,	HP*	739A- 1 6A	1		^	
						1

ΤÆ

* See "List of Manufacturers Code Letters For Replaceable Parts Table".

Total quantity used in the instrument.

LIST OF CODE LETTERS USED IN TABLE OF REPLACEABLE PARTS TO DESIGNATE THE MANUFACTURERS

,

LETTER	MANUFACTURER	ADDRESS	LETTER	MANUFACTURER	ADDRESS	
A	Aerovox Corp.	New Bedford, Mass.	AK	Hammerlund Mfg. Co., Inc.	New York I, N. Y.	
В	Allen-Bradley Co.	Milwaukee 4, Wis.	AL	Industrial Condenser Corp.	Chicago 18, Ill.	
С	Amperite Co.	New York, N.Y.	AM	Insuline Corp. of America	Manchester, N. H.	
D	Arrow, Hart & Hegeman	Hartford, Conn.	AN	Jennings Radio Mfg. Corp.	San Jose, Calif.	
E	Bussman Manufacturing Co.	St. Louis, Mo.	AO	E. F. Johnson Co.	Waseca, Minn.	
F	Carborundum Co	Niggara Falls, N. Y.	AP	Lenz Electric Mfg. Co.	Chicago 47. III	
G	Centralah	Milwaukee I. Wis	AQ	Micro-Switch	Freeport III	
ц	Circh Japas Mfr. Co	Chierge 24 III	AP	Machanical Industrias Prod. Co.	Akan & Ohio	
LID	Hawlett Backard Co	Bala Alta Calif	45	Medal Eng & Mfg Ing	Huntington Ind	
rir I	Character Mfa Co.	Palo Alto, Calif.	~J	The Muter Co	Chieners 5 III	
· ·	Clarostat Mig. Co.	Dover, N. H.		Oherita Miler Co.		
J	Cornell Dubilier Elec. Co.	South Plainfield, N. J.	AU	Onmite Mitg. Co.	Skokie, III.	
ĸ	HI-Q Division of Aerovox	Olean, N. T.	AV	Resistance Products Co.	Harrisburg, Pa.	
L	Erie Resistor Corp.	Erie 6, Pa.	AW	Kadio Condenser Co.	Camden 3, N. J.	
м	Fed. Telephone & Radio Corp.	Clifton, N. J.	AX	Shallcross Manutacturing Co.	Collingdale, Pa.	
N	General Electric Co.	Schenectady 5, N. Y.	AY	Solar Manufacturing Co.	Los Angeles 58, Calif.	
0	General Electric Supply Corp.	San Francisco, Calif.	AZ	Sealectro Corp.	New Rochelle, N. Y.	
Р	Girard-Hopkins	Oakland, Calif.	BA	Spencer Thermostat	Attleboro, Mass.	
Q	Industrial Products Co.	Danbury, Conn.	BC	Stevens Manufacturing Co.	Mansfield, Ohio	
R	International Resistance Co.	Philadelphia 8, Pa.	BD	Torrington Manufacturing Co.	Van Nuys, Calif.	
S	Lectrohm Inc.	Chicago 20, Ill.	BE	Vector Electronic Co.	Los Angeles 65, Calif.	
т	Littlefuse Inc.	Des Plaines, III.	BF	Weston Electrical Inst. Corp.	Newark 5, N. J.	
U	Maguire Industries Inc.	Greenwich, Conn.	BG	Advance Electric & Relay Co.	Burbank, Calif.	
V	Micamold Radio Corp.	Brooklyn 37. N. Y.	BH	E. I. DuPont	San Francisco, Calif.	
w	Oak Manufacturing Co.	Chicago 10, III.	BI	Electronics Tube Corp.	Philadelphia 18. Pa.	
x	P. R. Mallory Co., Inc.	Indianapolis, Ind.	BJ	Aircraft Radio Corp.	Boonton, N. J.	
Y	Radio Corp. of America	Harrison N.J.	BK	Allied Control Co., Inc.	New York 21, N.Y.	
7	Sanaamo Electric Co	Marion III	BL	Augat Brothers, Inc.	Attleboro Mass	
~ ~	Sarkes Tarzian	Bloomington Ind	BM	Carter Radio Division	Chicago III	
RR	Signal Indicator Co	Brooklyn 37 N Y	BN	CBS Hytron Radio & Electric	Danvars Mass	
CC	Saragua Electric Co	North Adams Mass	BO	Chicago Tolonhono Supply	Elkhart Ind	
DD	Strackpole Carbon Co	St Manue Pa	RP	Henry L Crowley Co. Inc.	West Orange N (
EE	Subartia Electric Products Co	Warron Pa	BO	Curtice Wright Com	Carlstadt N	
55	Sylvania Electric Froducts Co.	New York 5 N Y	RD	Allen B. Dublent Labe	Cliffor N. J.	
	Willes Products Inc	Claveland Ohio	BC	Excel Transformer Co	Critton, N. J.	
GG	wilkor Products, Inc.	Chierren E0 III	BT	General Padia Ca	Cambrida, Calif.	
	Amprenoi	Breaklyn 27 N. Y	BII		Cambridge 37, Mass.	
	Dial Light Co. of America	Brooklyn 37, N. T.	BV	Internet in all Partific Con	Culver City, Calif.	
JJ	Leecratt Manufacturing Co.	Chinese 22 III	D V	International Rectifier Corp.	El Segundo, Calif.	
KK	Switchcraft, Inc.		DYY	James Knights Co.	Sandwich, III.	
LL	Gremar Manufacturing Co.	Wakefield, Mass.	DA	Mueller Electric Co.	Cleveland, Ohio	
MM	Carad Corp.	Reawood City, Calif.	07	Precision Thermometer & Inst. Co.	Philadelphia 30, Pa.	
NN	Electra Manufacturing Co.	Kansas City, Mo.	BZ	Kadio Essentials Inc.	Mf. Vernon, N. Y.	
00	Acro Manufacturing Co.	Columbus 16, Ohio	CA	Raytheon Manufacturing Co.	Newton, Mass.	
PP	Alliance Manutacturing Co.	Alliance, Ohio	CB	lung-Sol Lamp Works, Inc.	Newark 4, N. J.	
QQ	Arco Electronics, Inc.	New York 13, N.Y.	CD	Varian Associates	Palo Alto, Calif.	
RR	Astron Corp.	East Newark, N. J.	CE	Victory Engineering Corp.	Union, N. J.	
SS	Axel Brothers Inc.	Long Island City, N.Y.	CF	Weckesser Co.	Chicago 30, Ill.	
TT	Belden Manufacturing Co.	Chicago 44, III.	CG	Wilco Corporation	Indianapolis, Ind.	
UU	Bird Electronics Corp.	Cleveland 14, Ohio	CH	Winchester Electronics, Inc.	Santa Monica, Calif.	
VV	Barber Colman Co.	Rockford, III.	CI	Malco Tool & Die	Los Angeles 42, Calif.	
WW	Bud Radio Inc.	Cleveland 3, Ohio	CJ	Oxford Electric Corp.	Chicago 15, Ill.	
XX	Allen D. Cardwell Mfg. Co.	Plainville, Conn.	CK	Camloc-Fastener Corp.	Paramus, N. J.	
YY	Cinema Engineering Co.	Burbank, Calif.	CL	George K. Garrett	Philadelph:a 34, Pa.	
ZZ	Any brand tube meeting		CM	Union Switch & Signal	Swissvale, Pa.	
	RETMA standards.		CN	Radio Receptor	New York II, N. Y.	
AB	Corning Glass Works	Corning, N.Y.	CO	Automatic & Precision Mfg. Co.	Yonkers, N. Y.	
AC	Dale Products, Inc.	Columbus, Neb.	CP	Bassick Co.	Bridgeport 2, Conn.	
AD	The Drake Mfg. Co.	Chicago 22, Ill.	CQ	Birnbach Radio Co.	New York 13, N.Y.	
AE	Elco Corp.	Philadelphia 24, Pa.	CR	Fischer Specialties	Cincinnati 6, Ohio	
AF	Hugh H. Eby Co.	Philadelphia 44, Pa.	CS	Telefunken (c/o MVM, Inc.)	New York, N.Y.	
AG	Thomas A. Edison, Inc.	West Orange, N. J.	CT	Potter-Brumfield Co.	Princeton, Ind.	
AH	Fansteel Metallurgical Corp.	North Chicago, Ill.	CU	Cannon Electric Co.	Los Angeles, Calif.	
AI	General Ceramics & Steatite Corp.	Keasbey, N. J.	CV	Dynac, Inc.	Palo Alto, Calif.	
AJ	The Gudeman Co.	Sunnyvale, Calif.	CW	Good-All Electric Mfg. Co.	Ogallala, Nebr.	

CLAIM FOR DAMAGE IN SHIPMENT

The instrument should be tested as soon as it is received. If it fails to operate properly, or is damaged in any way, a claim should be filed with the carrier. A full report of the damage should be obtained by the claim agent, and this report should be forwarded to us. We will then advise you of the disposition to be made of the equipment and arrange for repair or replacement. Include model number and serial number when referring to this instrument for any reason.

6,-

WARRANTY

Hewlett-Packard Company warrants each instrument manufactured by them to be free from defects in material and workmanship. Our liability under this warranty is limited to servicing or adjusting any instrument returned to the factory for that purpose and to replace any defective parts thereof. Klystron tubes as well as other electron tubes, fuses and batteries are specifically excluded from any liability. This warranty is effective for one year after delivery to the original purchaser when the instrument is returned, transportation charges prepaid by the original purchaser, and when upon our examination it is disclosed to our satisfaction to be defective. If the fault has been caused by misuse or abnormal conditions of operation, repairs will be billed at cost. In this case, an estimate will be submitted before the work is started.

If any fault develops, the following steps should be taken:

1. Notify us, giving full details of the difficulty, and include the model number and serial number. On receipt of this information, we will give you service data or shipping instructions.

2. On receipt of shipping instructions, forward the instrument prepaid, to the factory or to the authorized repair station indicated on the instructions. If requested, an estimate of the charges will be made before the work begins provided the instrument is not covered by the warranty.

SHIPPING

All shipments of Hewlett-Packard instruments should be made via Truck or Railway Express. The instruments should be packed in a strong exterior container and surrounded by two or three inches of excelsior or similar shock-absorbing material.

DO NOT HESITATE TO CALL ON US

HEWLETT-PACKARD COMPANY Laboratory Instruments for Speed and Accuracy 275 PAGE MILL ROAD CABLE PALO ALTO.CALIF. U.S.A. "HEWPACK"